

# ARCS Instrument Replacement Form

Fill out the following form when replacing instruments at an ARCS site:

SV-0611\_repair

System: Skyrad NIP

Site: Darwin

## OLD Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) $\times 10^{-6}$ or 4 PIR coefs.	Offset	Date of Change (UTC)	Time Start UTC	Time End UTC
NIP	31351E6	Skyrad Tracker	9/11/2006	8.2521	na	11/20/2005	06:10	06:36

## NEW Configuration

Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45°) $\times 10^{-6}$ or 4 PIR coefs.	Offset	Change Logger Coefficients and Version
NIP	31350E6	Skyrad Tracker	11/20/2006	8.3620	na	X

Replacement completed by: Troy Culgan

## NOTES:

Old Part remains on site as spare:

Old part packed for shipment to: SGP for repair

Other old part disposition (explain):

Other comments: NIP replaced due to erroneously high readings